

FOX CHASE CANCER CENTER

Department of Radiation Oncology
7701 Burholme Avenue
Philadelphia, Pennsylvania 19111
215 728 2581

October 17, 2000

Krishnan Suthanthiran, Meng
President
Best Medical International
7643 Fullerton Rd
Springfield, VA 22153


Dear Krish,

As per our conversation earlier last week, I just completed a review of our implants using the Best Model 2301 Iodine-125 seeds done recently at our institution. We have done eight permanent prostate implants in the last few weeks using Best Model 2301 Iodine-125 seeds and the results have been very good. The policy of our institution is to assay 5% of all radioactive seeds and the assay of approximately 900 seeds used recently shows a deviation of less than 2% from the activity stated on the calibration certificate.

We have also used the new spacers from Best Industries for the last couple of implants. We use the SeedVac loading system from Standard Imaging and the quality of the spacers becomes an important factor. The manufacturing quality of your spacers is very impressive and the ease of loading the needles is greatly improved. I am enclosing a couple of radiographs showing the high quality of prostate implant achieved using the products from Best Industries.

I hope this provides you with some feedback on your recently introduced products. If there are any questions, please contact me directly.

Best Regards,



Raj K. Mitra, M.S., DABMP